

***E. coli* (Protein Expression Strains) HCP ELISA Kit
(One-step ELISA)**

Product No.: **1301301-1**

Lot No.: **20250502**

Date of Mfg.: **29 May 2025**

Expiration Date: **23 Apr. 2026**

Test Item	Specification	Result
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	81 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Limit of Quantification	1.5 ng/mL sample solution was tested, $CV \leq 20.0\%$, relative bias of concentration was not more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS (OD _{450 nm}) should be not more than 0.200.	PASS

Approved By Zong Meiyi

DAY/MONTH/YEAR 03/06/2025
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.com

For any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China